

Sensor metadata

Type: Differential pore pressure and temperature sensors

2017

Model : Piezometer V2

Provider: NKE

Serial number

PT1 @ 0.79m	1145
PT2 @ 1.59m	1153
PT3 @ 2.39m	1154
PT4 @ 3.19m	1155
PT5 @ 3.99m	1156
PT6 @ 4.79m	1164
PT7 @ 5.59m	1165

Operating parameters

Calibration: before deployment, 4-points calibration and zero checking

Sampling rate: 2 minutes